

.

From: Khan, Annie (DPH)
Sent: Thursday, December 22, 2005 10:05 AM
To: Salemi, Charles (DPH)
Subject: GC # 2

Hi Chuck,

GC # 2 is not working well with the coke/codeine std. I ran 3 individual vials of the standard and they did not look like my QC from Monday.

The Cocaine is fine, but the Codeine is not good. I left the chromatograms in your box.

-Annie